

## SCANNING ELECTRON MICROSCOPY SCHEME (SEMS)

Laboratory No.	2023	Name Of Laboratory	Tampereen Asbesti - Ja Kuitulaboratorio
Round No.	008A	Laboratory Representative	Heli Knuutila
		E-Mail Address	tampereen@asbestilaboratorio.fi
Report No.	0179	Report Issued	Nov 7 2018 11:20AM

SAMPLE	No	Amphibole Density	Chrysotile Density	Inorganic Density	Total Asbestos	Median	Band
8A SEMS 1	1	0.00	0.00	35.10	0.00	0	A
	2	0.00	0.00	51.20	0.00	0	A
	3	0.00	0.00	72.10	0.00	0	A
8A SEMS 2	1	12.30	0.00	0.00	12.30	17.1	A
	2	17.10	0.00	0.00	17.10	17.1	A
	3	22.80	0.00	0.00	22.80	17.1	A
8A SEMS 3	1	30.40	0.00	0.00	30.40	29	A
	2	37.90	0.00	0.00	37.90	29	A
	3	38.90	0.00	0.00	38.90	29	A
8A SEMS 4	1	0.00	16.20	0.00	16.20	16.4	A
	2	0.00	27.50	0.00	27.50	16.4	A
	3	0.00	22.80	0.00	22.80	16.4	A

### Total Number In Each Band

-C	-B	A	B	C
0	0	12	0	0

Details of performance assesement are given in  
'SEMS Information For Participants'.

Number Of Valid Results	12	
Results Within Band A	12	100%
Results Within Band A + B	12	100%

Melanie Clunas  
SEM Scheme Co-Ordinator  
Proficiency Testing Team



Health & Safety Laboratory  
Harpur Hill, Buxton, Derbyshire SK17 9JN - UK  
Report No. : 0179