

# CENTRIFUGE SETTINGS



Regen Lab only guarantees RegenBCT, RegenATS and RegenTHT tubes' performance when centrifuged with an original Regen Lab device, according to the instruction for use manual.

		
1500g = 31x100RPM	1500g = 31x100RPM	1500g = 34x100RPM



**Tubes should be properly balanced in the centrifuge.**

If needed use a counterbalance tube filled with the equivalent volume of water.

